

**CALIBRADORES DE ROSCAS METRICAS - TAPON ROSCADO -
NORMA DIN 13 ISO 1502 ROSCA INTERIOR Calidad 6 H**



Medida Nominal		Medidas Generales				Medidas Nominales			Tolerancias de Fabricacion				
Ø	Paso	LT	LP	LNP	PASO	Ø Primitivo		Ø Exterior	Ø Primitivo 0 / +		Ø Exterior 0/+	Paso +/1	Semiang. < + / -
						pasa	no pasa		pasa	no pasa	pasa		
"	h x 1"	mm	mm	mm	mm	µ	µ	µ	µ	µ	µ	µ	'
2.0	0,4	77	4,00	4,00	0,40	1,74	1,83	2,000	1,5 / 10 5	9	- 3 / + 15	5	28°
3.0	0,45	81	8,00	4,00	0,45	2,208	2,303	2,500	1,5 / 10 6	9	- 3 / + 15	5	26°
3.5	0,50	81	8,00	4,00	0,50	2,675	2,775	3,000	1,5 / 10 7	9	- 3 / + 15	5	24°
4.0	0,60	83	9,00	4,5	0,60	3,11	3,222	3,500	1,5 / 10 8	9	- 3 / + 15	5	21°
4.0	0,50	83	9,00	4,5	0,50	3,675	3,775	4,000	1,5 / 10 9	9	- 3 / + 15	5	24°
5.0	0,70	83	9,00	4,5	0,70	3,545	3,663	4,000	1,5 / 10 10	9	- 3 / + 15	5	18°
5.0	0,50	85	10,00	5,0	0,50	4,675	4,775	5,000	1,5 / 10 11	9	- 3 / + 15	5	24°
6.0	0,80	85	10,00	5,0	0,80	4,48	4,605	5,000	1,5 / 10 12	9	- 3 / + 15	5	16°
6.0	0,75	95	12,5	6,3	0,75	5,513	5,645	6,000	6,5 / 17,5	11	+1 / +23	5	17°
7.0	1,00	95	12,5	6,3	1,00	5,35	5,5	6,000	6,5 / 17,5	11	+1 / +23	5	14°
8.0	1,00	95	12,5	6,3	1,00	6,35	6,5	7,000	6,5 / 17,5	11	+1 / +23	5	14°
8.0	0,75	95	12,5	6,3	0,75	7,513	7,645	8,000	6,5 / 17,5	11	+1 / +23	5	17°
8.0	1,00	95	12,5	6,3	1,00	7,35	7,5	8,000	6,5 / 17,5	11	+1 / +23	5	14°
9.0	1,25	95	12,5	6,3	1,25	7,188	7,348	8,000	6,5 / 17,5	11	+1 / +23	5	13°
9.0	1,00	95	12,5	6,3	1,00	8,35	8,5	9,000	6,5 / 17,5	11	+1 / +23	5	14°
10.0	1,25	95	12,5	6,3	1,25	9,35	9,348	9,000	6,5 / 17,5	11	+1 / +23	5	13°
10.0	1,00	104	10,0	6,3	1,00	9,188	9,500	10,000	6,5 / 17,5	11	+1 / +23	5	14°
10.0	1,25	110	16,0	6,3	1,25	9,026	9,348	10,000	6,5 / 17,5	11	+1 / +23	5	13°
12.0	1,00	110	16,0	6,3	1,50	11,35	9,206	10,000	6,5 / 17,5	11	+1 / +23	5	12°
12.0	1,25	104	10,0	6,3	1,00	13,188	11,510	12,000	6,5 / 17,5	11	+1 / +23	5	14°
12.0	1,50	110	16,0	6,3	1,25	13,026	11,638	12,000	6,5 / 17,5	11	+1 / +23	5	13°
12.0	1,00	110	16,0	6,3	1,50	12,701	11,216	12,000	6,5 / 17,5	11	+1 / +23	5	12°
14.0	1,25	114	16,0	10,0	1,75	15,35	11,063	12,000	6,5 / 17,5	11	+1 / +23	5	11°
14.0	1,50	121	12,5	8,0	1,00	13,188	13,510	14,000	6,5 / 17,5	11	+1 / +23	5	14°
14.0	1,75	121	12,5	8,0	1,25	13,026	1,368	14,000	6,5 / 17,5	11	+1 / +23	5	13°
14.0	1,00	129	20,0	8,0	1,50	12,701	13,216	14,000	6,5 / 17,5	11	+1 / +23	5	12°
16.0	1,75	129	20,0	8,0	2,00	15,350	12,913	14,000	9,0 / 23,0	14	+2 / +30	5	10°
16.0	1,00	121	12,5	8,0	1,00	15,026	15,510	16,000	6,5 / 17,5	11	+1 / +23	5	14°
16.0	1,25	129	20,0	8,0	1,50	14,701	15,216	16,000	6,5 / 17,5	11	+1 / +23	5	12°
16.0	1,50	129	20,0	8,0	2,00	17,026	14,913	16,000	9,0 / 23,0	14	+2 / +30	5	10°
18.0	2,00	140	16,0	10,0	1,50	16,376	17,216	18,000	6,5 / 17,5	11	+1 / +23	5	12°
18.0	1,00	140	25,0	10,0	2,50	17,026	16,600	18,000	9,0 / 23,0	14	+2 / +30	5	10°